



**KARNATAKA ANTIBIOTICS &
PHARMACEUTICALS LIMITED**

(A Government of India Enterprise)

ENQUIRY REF. No.	KAPL/QAD/020/0802
DATE	05/07/2024
DUE DATE	10/07/2024 (13.00HRS)

Dear Sir,

Please submit your lowest and competitive offer in a SEALED ENVELOPE, DULY SUPERSCRIBING OUR ABOVE ENQUIRY REF. NO., DATE and DUE DATE on it/ OR MAIL, with other details of F.O.R terms, Taxes, Credit period, Delivery offered, Name of the Make, Detailed Specification etc., for below mentioned material/s

SL. NO.	ITEM CODE	ITEM DESCRIPTION	UOM	QTY
01	QSPHPL374	15CMX4.6MM 5µM ODS POROUS SILICA	NOS	01


OTHER TERMS:

- | | |
|---------------------------------|------------------|
| 1. F.O.R TERMS | : DOOR DELIVERY |
| 2. GST % | : PLEASE SPECIFY |
| 3. PACKING & FORWARDING CHARGES | : NOT APPLICABLE |
| 4. CREDIT PERIOD | : 30 DAYS |
| 5. DELIVERY OFFERED | : |

NOTE: IN CASE YOU ARE NOT QUOTING PLEASE SEND THE REGRET LETTER.

Thanking you,

Yours faithfully,
For KARNATAKA ANTIBIOTICS
& PHARMACEUTICALS LIMITED


YUVARAJA M
DEPUTY MANAGER PURCHASE DEPT



KARNATAKA ANTIBIOTICS AND PHARMACEUTICALS LIMITED, BENGALURU

CEPHALOSPORIN QUALITY CONTROL DEPARTMENT

User Requirement Specifications

Material Description: HPLC C18 Column 80 Å, 150mm x 4.6 mm x 5 µm (Spherisorb)

URS Number: CEPHAQC/URS/003

1. Description and Quantity

Material Description	HPLC C18 Column 80 Å, 150mm x 4.6 mm x 5 µm (Spherisorb)
Item Code	QSPHPL374
Quantity/Box	1 Nos

2. User Specifications

#	Requirement	Specification
1.	Brand Name	HPLC C18 Column 80 Å, 150mm x 4.6 mm x 5 µm (Spherisorb)
2.	Make	WATERS
3.	Part No.	PSS830613
4.	Particle size	5 µm
5.	Surface Area (m ² /g)	220
6.	Length	150mm
7.	Internal Diameter (I.D.)	4.6 mm
8.	Pore Size	80 Å
9.	Carbon load	6%
10.	Particle Substrate	Silica
11.	Endcapped	No
12.	Mode of Chromatography	Reversed Phase
13.	Construction Materials	Stainless Steel
14.	USP Classification	L1
15.	Separation Mode	Reversed Phase
16.	pH Range	2-8
17.	Maximum Pressure	6000 psi (415 Bar)