



ENQUIRY REF. No.	KAPL/QAD/020/0844
DATE	11/07/2024
DUE DATE	16/07/2024 (13.00HRS)

Dear Sir,

Please submit your lowest and competitive offer in a SEALED ENVELOPE, DULY SUPERSCRIBING OUR ABOVE ENQUIRY REF. NO., DATE and DUE DATE on it/ OR MAIL, with other details of F.O.R terms, Taxes, Credit period, Delivery offered, Name of the Make, Detailed Specification etc., for below mentioned material/s

SL. NO.	ITEM CODE	ITEM DESCRIPTION	UOM	QTY
01	QSPHPL417	HPLC COLUMN 15CM X 4.6MM,C8,5 μ (SPHERISORB ODS-1)	NOS	01
02	QSPHPL422	HPLC COLUMN 15CM X 3.9MM,C8,5 μ (SYMMETRY C8)	NOS	01
03	QSPHPL416	HPLC COLUMN 15CM X 4.6MM,ODS,5 μ (SPHERISORB ODS-2)	NOS	01

OTHER TERMS:

- | | |
|---------------------------------|------------------|
| 1. F.O.R TERMS | : DOOR DELIVERY |
| 2. GST % | : PLEASE SPECIFY |
| 3. PACKING & FORWARDING CHARGES | : NOT APPLICABLE |
| 4. CREDIT PERIOD | : 30 DAYS |
| 5. DELIVERY OFFERED | : |

NOTE: IN CASE YOU ARE NOT QUOTING PLEASE SEND THE REGRET LETTER.

Thanking you,

Yours faithfully,
For KARNATAKA ANTIBIOTICS
& PHARMACEUTICALS LIMITED


YUVARAJA M
DEPUTY MANAGER PURCHASE DEPT

**User Requirement specifications****Material Description** : HPLC COLUMN 15cm x 4.6mm, C8, 5u**URS Number**: QC/URS/004/0724**1. Description and Quantity:**

Material Description	25cm x 4.6mm, C8, 5u
Item code	QSPHPL417
Quantity/ Box	1

2. User Specifications:


#	Requirement	Specification
1	Brand Name	15cm x 4.6mm, C8, 5u (SPHERISORB ODS-1)
2	Make	Waters
3	Brand	Spherisorb
4	Part Number	PSS830613
5	Matrix active group	Silica
6	Particle size	5u
7	Length (mm)	150
8	Internal Diameter (I.D.)	4.6 mm
9	Particle Substrate	Silica
10	Particle Shape	Spherical
11	External Construction Materials	Stainless Steel
12	Endcapped	No
13	Endfitting Type	Parker-style
14	USP Classification	L1
15	Separation Mode	Reverse phase
16	P ^H Range	2-8
17	Maximum Pressure	6000 psi (415 Bar)
18	Pore Size	80 °A

**User Requirement specifications****Material Description** : HPLC COLUMN 15cm x 3.9mm, C8, 5u**URS Number**: QC/URS/005/0724**1. Description and Quantity:**

Material Description	15cm x 3.9mm, C8, 5u
Item code	QSPHPL422
Quantity/ Box	1

2. User Specifications:

#	Requirement	Specification
1	Brand Name	15cm x 3.9mm, C8, 5u (Symmetry C8)
2	Make	Waters
3	Brand	Symmetry
4	Part Number	WAT046970
5	Matrix active group	Silica
6	Particle size	5u
7	Length (mm)	150
8	Internal Diameter (I.D.)	3.9 mm
9	Particle Substrate	Silica
10	Particle Shape	Spherical
11	External Construction Materials	Stainless Steel
12	Endcapped	Yes
13	Endfitting Type	Waters
14	USP Classification	L7
15	Separation Mode	Reverse phase
16	P ^H Range	2-8
17	Maximum Pressure	6000 psi (415 Bar)
18	Pore Size	100 °A

QUALITY CONTROL DEPARTMENT		KARNATAKA ANTIBIOTICS & PHARMACEUTICALS LIMITED
		(A Government of India Enterprise)

User Requirement specifications

Material Description : HPLC COLUMN 15cm x 4.6mm, ODS, 5u

URS Number: QC/URS/006/0724

1. Description and Quantity:

Material Description	15cm x 4.6mm, ODS, 5u
Item code	QSPHPL416
Quantity/ Box	1

2. User Specifications:

#	Requirement	Specification
1	Brand Name	15cm x 4.6mm, ODS, 5u (Spherisorb ODS-2)
2	Make	Waters
3	Brand	Spherisorb
4	Part Number	PSS831913
5	Matrix active group	Silica
6	Particle size	5u
7	Length (mm)	150
8	Internal Diameter (I.D.)	4.6 mm
9	Particle Substrate	Silica
10	Particle Shape	Spherical
11	External Construction Materials	Stainless Steel
12	Endcapped	Yes
13	Endfitting Type	Parker-style
14	USP Classification	L1
15	Separation Mode	Reverse phase
16	P ^H Range	2-8
17	Maximum Pressure	6000 psi (415 Bar)
18	Pore Size	80 °A