



**KARNATAKA ANTIBIOTICS &
PHARMACEUTICALS LIMITED**

(A Government of India Enterprise)

ENQUIRY REF. No.	KAPL/CG/011/1329
DATE	13/09/2024
DUE DATE	23/09/2024 (13.00 HRS)

Dear Sir,

Please submit your lowest and competitive offer in a SEALED ENVELOPE, DULY SUPERSCRIBING OUR ABOVE ENQUIRY REF. NO., DATE and DUE DATE on it. Mentioned other details of F.O.R terms, Taxes, Credit period, Delivery offered, Name of the Make, Detailed Specification etc., for the below Supply or Service

SI No	ItemCode	Item Description	Uom	Qty
01	TNP2024PRT	WEIGHING BALANCE 600G,10MG ACCURACY WITH PRINTER	NOS	01

Please ensure that your offer reaches us on or before Due Date by courier OR Speed post or by Hand in sealed cover only to below office address:

*M/s. Karnataka Antibiotics and Pharmaceuticals Limited
Plot No.37, Arka The Business Centre,NTTF Main Road, Peenya Industrial Area
2nd Phase, Bengaluru-560058 Ph. No.080-23571590*

OTHER TERMS:

- | | |
|---------------------------------|------------------|
| 1. F.O.R TERMS | : DOOR DELIVERY |
| 2. GST % | : PLEASE SPECIFY |
| 3. PACKING & FORWARDING CHARGES | : NOT APPLICABLE |
| 4. CREDIT PERIOD | : 30 DAYS |
| 5. DELIVERY OFFERED | : PLEASE SPECI |

NOTE:

01. IF YOU ARE NOT PARTICIPATING IN THE TENDER PLEASE SEND A REGRET LETTER.

02. VENDOR HAS TO QUOTE AS PER OUR TENDER IN YOUR COMPANY LETTER HEAD.

03. **QUOTATION MUST BE SUBMITTED IN TWO SEALED COVERS (TECHNICAL & COMMERCIAL/PRICE BID) SEPARATELY AND IN ONE ENVELOPE OR ELSE YOUR PROPOSAL WILL NOT BE CONSIDERED.**

IF YOU NEED ANY CLARIFICATION, PLEASE CONTACT US.

Thanking you,

Yours faithfully,
For KARNATAKA ANTIBIOTICS
& PHARMACEUTICALS LIMITED



YUVARAJA.
DY. MANAGER – PURCHASE DEPT.
MOB: 9945317873



**USER REQUIREMENT
SPECIFICATIONS
FOR
WEIGHING BALANCE**

Tha
You
For I
& PH
YUVA
DY. M.
MOB: 5

**Karnataka Antibiotics & Pharmaceuticals Limited
NON PARENTERALS DEPARTMENT**

User Requirement Specifications



Equipment/system	WEIGHING BALANCE
Document#	NP/URS/WB/01

2. Objective

This document specifies the list of user requirements for weighing balance.

3. Purpose

The purpose of the instrument is for weighing powder, tablets and capsules at non parenterals department.

4. Scope

The scope of the equipment is bound to this URS and any changes will be through revised and approved URS. Vendor should provide a demonstration of his latest model before quoting

Description	Number of Units required
Digital weighing balance with printer	4 Units

1. User Specifications

Specification	Range
1. Weighing Capacity	20mg -620g
2. Decimal	3 decimals
3. e	10mg
4. d	d=0.001g / 1mg
5. Weighing pan dimensions (mm)	90mm
6. Warning	min-weigh warning function, Strong overload protection
7. Other Features;	CAL functions for automatic internal calibration with variation in temperature or after elapsing a time interval that has been set.
	Built in motorized calibration weight for internal calibration Function.

PRINTER Necessary & Shall be suitable for Pharma Laboratory

- a) Print out shall contains:
- b) Name of the company
- c) Date
- d) Start time
- e) Name of the Balance
- f) Model
- g) Sr.No.
- h) Ver.No.
- i) Eqp. ID:
- j) Batch No.
- k) Details of weight : w1,w2.....,Total, Avg, SD, RSD,
- l) Min. wt, Max. wt, Range (Diff) etc.

Equipment/system

WEIGHING BALANCE

Document#

NP/URS/WB/01

**8 Scope of Training**

- The vendor/supplier shall provide training at site at no extra cost. This shall include, but not limited to, making available the services of a qualified representative to train the operators.
- The training of the personnel shall include a minimum of the following to the complete satisfaction of KAPL team:

Start-up and shut-down procedures.

Typical operations

Emergency shutdown

Safety training

Troubleshooting procedure.

- The vendor/supplier shall ensure that the operation manual is available at site for reference. The vendor/supplier shall issue certificate to the persons trained by them after successful completion of training at no extra cost.

9 Installation and commissioning

The vendor/supplier shall ensure the correct positioning, to the satisfaction of KAPL and be responsible for the installation of the equipment, IQ/OQ and PQ.

The vendor /supplier shall also be responsible for the re-assembly of sections or components of the packaged equipment, if any including wiring, instrumentation and any other equipment and components shipped separately or in sections.

Than

Yours
For K,
& PHAYUVAR
DY. MA.
MOB: 99

Karnataka Antibiotics & Pharmaceuticals Limited
NON PARENTERALS DEPARTMENT

User Requirement Specifications

Equipment/system	WEIGHING BALANCE
Document#	NP/URS/WB/01



- | | |
|----|------------|
| m) | End time: |
| n) | Name: |
| o) | Signature: |

Miscellaneous

- | | |
|----|--|
| a) | Display: It should show all critical parameters. |
| b) | Wind shield :- polycarbonate/glass |

Service, Warranty and Training

- | | |
|----|--|
| 1. | Training (all Users) at Customers location |
| 2. | On-site installation, commissioning and training shall be conducted by a qualified factory-trained engineer. |
| 3. | Warranty for equipment for a period of 24 months. |
| 4. | Vendor to provide both on-site and training for users on the system start-up, usage , maintenance, trouble shooting, etc |

Others

- | | |
|----|--|
| 1. | Screen operate with gloves(cotton, silicon or rubber) |
| 2. | Chemical -Resistant full metal housing. |
| 3. | proFact internal adjustment |
| 4. | ISO Log, Sample ID input, Passcode Protection. |
| 5. | Quick to dismantle, QuickLock for glass panels, QuickLock for draft shield. |
| 6. | Vendor must invite KAPL team member/s for a demo of similar equipment at their facility. |

5 Scope of installation

Training must be provided by the vendor, to the satisfaction of KAPL team.

6 Warranty

Minimum 1 year warranty or as per PO terms and conditions.

7 Documentation

Documentation Supplier shall provide the following in English language.

- Operational and maintenance manual.
- Operation manual.
- Test Certificates
- Installation and operational protocols & Reports.
- Calibration certificates.
- Engineer qualified certificate