



**KARNATAKA ANTIBIOTICS &
PHARMACEUTICALS LIMITED**

(A Government of India Enterprise)

ENQUIRY REF. No.	KAPL/CG/011/1344
DATE	13/09/2024
DUE DATE	23/09/2024

Dear Sir,

Please submit your lowest and competitive offer in a SEALED ENVELOPE, DULY SUPERSCRIBING OUR ABOVE ENQUIRY REF. NO., DATE and DUE DATE on it. Mentioned other details of F.O.R terms, Taxes, Credit period, Delivery offered, Name of the Make, Detailed Specification etc., for the below Supply or Service

SI No	ItemCode	Item Description	Uom	Qty
01	TSVP1028	INSPECTION TABLE FOR AMBER COLOUR AMPOULES /VIALS	NOS	02

Please ensure that your offer reaches us on or before Due Date by courier OR Speed post or by Hand in sealed cover only to below office address:

*M/s. Karnataka Antibiotics and Pharmaceuticals Limited
Plot No.37, Arka The Business Centre,NTTF Main Road, Peenya Industrial Area
2nd Phase, Bengaluru-560058 Ph. No.080-23571590*

TECHNICAL CLARIFICATION

1) URS ATTACHED 2 PAGES

OTHER TERMS:

- | | |
|---------------------------------|------------------|
| 1. F.O.R TERMS | : DOOR DELIVERY |
| 2. GST % | : PLEASE SPECIFY |
| 3. PACKING & FORWARDING CHARGES | : NOT APPLICABLE |
| 4. CREDIT PERIOD | : 30 DAYS |
| 5. DELIVERY OFFERED | : PLEASE SPECIFY |

NOTE:

01. IF YOU ARE NOT PARTICIPATING IN THE TENDER PLEASE SEND A REGRET LETTER.
02. VENDOR HAS TO QUOTE AS PER OUR TENDER IN YOUR COMPANY LETTER HEAD.

03. QUOTATION MUST BE SUBMITTED IN TWO SEALED COVERS (TECHNICAL & COMMERCIAL/PRICE BID) SEPARATELY AND IN ONE ENVELOPE OR ELSE YOUR PROPOSAL WILL NOT BE CONSIDERED.

IF YOU NEED ANY CLARIFICATION, PLEASE CONTACT US.

Thanking you,

Yours faithfully,
For KARNATAKA ANTIBIOTICS
& PHARMACEUTICALS LIMITED

YUVARAJA. M
DY. MANAGER – PURCHASE DEPT.
MOB: 9945317873

Karnataka Antibiotics & Pharmaceuticals Limited
ENGINEERING DEPARTMENT

User Requirement Specifications

Equipment/system	INSPECTION TABLE FOR AMBER COLOURED AMPOULES / VIALS		
Identification #	ITAV/01	Document#	SVP/URS/ITAV/01
Effective Date		Revision#	00



**USER REQUIREMENT SPECIFICATIONS
FOR
INSPECTION TABLE FOR AMBER COLOURED AMPOULES /
VIALS**

Revision index:

Revision	Date	Reason for revision
00	12.09.2024	Initial Document

Annexure List:

Annex No.	Drawing / Document No.	Detail
1		

Location of Installation:

In small volume parenteral block

Room Name: VISUAL INSPECTION ROOM

Karnataka Antibiotics & Pharmaceuticals Limited

ENGINEERING DEPARTMENT

User Requirement Specifications

Equipment/system	INSPECTION TABLE FOR AMBER COLOURED AMPOULES / VIALS		
Identification #	ITAV/01	Document#	SVP/URS/ITAV/01
Effective Date		Revision#	00



1. OBJECTIVE:

Since parenteral preparations are injected directly into the human body, it is very much essential that they are completely free from particulate matter. Following procedure provides guidelines for visual inspection of vials/ampoules through which we should be able to produce particle free vials & ampoules.

2. SCOPE:

This URS is applicable to inspection table for visual inspection of amber coloured vials/ampoules for particulate matter and other quality defects.

3. SPECIFICATIONS:

NOMENCLATURE:

Equipment TAG Number	ITAV / 01
Qty.	02 No's

CONSTRUCTION DETAILS OF MACHINE:

Dimensions	1550mm(L) X 725mm (W) X 1245mm (H)
M.O.C.	SS304 with Smooth finish for corners and welded joints
Inspection Type	Manual Inspection Table against black & white background under light (04 Lights)
Model	GMP
Both Side Working Space	356mm
No. of operators	4 nos. (2 x 2 at both side – 4 persons)
Working Height	762mm
Background	White and Black with Matte or Non Glossy backgrounds are recommended in order to avoid interference from reflection
Dimension of Background	445mm (H) X 368mm (L)
No. of Backgrounds	08 nos. i.e. 2 White and 2 Black backgrounds at one side. (Alternate white and black Background for 4Persons for one table)
Light Lux Requirements	Light lux at the viewing point should be between 8,000 to 10,000 Lux for amber coloured ampoules / vials
Electrical Supply	230 VAC

DOCUMENTS & QUALIFICATION REQUIREMENTS:

1. Calibration and warranty certificates
2. Test Certificates of all material of construction
3. List of all components with referencing
4. Pre installation requirements
5. Installation requirements
6. DQ, IQ, PQ documents
7. Any clarifications should be cleared before quoting & no deviation from the URS permitted after quoting