

KARNATAKA ANTIBIOTICS & PHARMACEUTICALS LIMITED

(A Government of India Enterprise)

ENQUIRY REF. No.	KAPL/FD/SR/1369/2024-25
DATE	19,09.2024
DUE DATE	24.09.2024 till 1:00 PM

Dear Sir,

Please submit your lowest and competitive offer in a SEALED ENVOLOPE, DULY SUPERSCRIBING OUR ABOVE ENQUIRY REF. NO., DATE and DUE DATE on it. Mentioned other details of F.O.R terms, Taxes, Credit period, Delivery offered, Name of the Make, Detailed Specification etc., for the below Supply or Service

SI	Item Code	Item Description	Uom	Qty
NO		CALIBRATION OF E2 WEIGHT RANGING FROM IMG TO 200 GRAMS DETAILS AS PER ANNEXURE ENCLOSED	Nos.	23

Please ensure that your offer reaches us on or before Due Date by courier OR Speed post or by Hand in sealed cover to below office address or email to purenp@gmail.com:

M/s. Karnataka Antibiotics and Pharmaceuticals Limited Plot No.37, Arka The Business Centre, NTTF Main Road, Peenya Industrial Area 2nd Phase, Bengaluru-560058 Ph. No.080-23571590

NOTE:

- 01. PLEASE VISIT OUR SITE FOR BETTER UNDERSTANDING THE SCOPE OF WORK
- 02. IF YOU ARE NOT PARTICIPATING IN THE TENDER PLEASE SEND A REGRET LETTER.
- 03. VENDOR HAS TO QUOTE AS PER OUR TENDER IN YOUR COMPANY LETTER HEAD.

IF YOU NEED ANY CLARIFICATION, PLEASE CONTACT US.

Thanking you,

Yours faithfully,

For KARNATAKA ANTIBIOTICS & PHARMACEUTICALS LIMITED

VAVARAJA. M

DY. MANAGER - PURCHASE DEPT.

MOB: 9945317873



KARNATAKA ANTIBIOTICS AND PHARMACEUTICALS LIMITED, BENGALURU

RESEARCH AND DEVELOPMENT

User Requirement Specifications

Service Description: Calibration of E2 Weight Ranging from 1 mg to 200 g

URS Number: RD/URS/001/0924/V1

1. Description

Material Description	Calibration of E2 Weight Ranging from 1 mg to 200 g	
Item Code	SCALFD001	

2. User Specifications

1.	Number of single weights	Total: 23 1 mg(1 unit), 2 mg(2 units), 5 mg(1 unit), 10 mg(1 unit), 20 mg(2 units), 50 mg(1 unit), 100 mg(1 unit), 200 mg(2 units), 500 mg(1 unit), 1 g(1 unit), 2 g(2units), 5 g(1 unit), 10 g(1 unit), 20 g(2 units), 50 g(1 unit), 100 g(1 unit), 200 g(2 units)	
2.	Accuracy Class of Weight	1 mg to 200 g OIML Precision E2 Class Weights	
3.	Calibration Validity	NLT 2 Years from date of Calibration	

3. Documentation:

Supplier must provide the valid Calibration certificates.

Prepared by	Checked by	Approved by
Pah'L 17/09/12024	Pauray 17/09/2024	THE OFFICE
Analyst-R&D Akshay	AM-R&D Pushpendu Gaurav	DGM-R&D Panneer Selvam R